

# **Microbiological Laboratory Quality Assurance**

## Laboratory quality assurance

*A group of activities proceeded in order to ensure the validity of test results*

# Laboratory quality assurance according to ISO/IEC17025 clause 5.9

1. Internal quality control
2. External quality assessment

# 1. Internal quality control

**Internal quality control** consists of all the procedures undertaken by a laboratory for the continuous evaluation of its work in order to ensure the consistency of results day-to-day and their conformity with defined criteria

## 1. Internal quality control

### 1.1 Personel

- On the job training
- Competency test
- Job assignment

## 1. Internal quality control

### 1.2 Environment

- Room temperature
- Air ventilation
- Air and work area cleanliness
- Humidity

## 1. Internal quality control

### 1.3 Equipment

- Equipment profile
  - installation
  - Repair and maintenance
- Manual and instruction
- Preventive maintenance program
- Calibration
- Performance check

## 1. Internal quality control

### 1.4 Reference culture

- Growth performance of culture media
- Method validation
- Quality control of test
- Competency test

## 1. Internal quality control

### 1.5 Culture media quality and water supplied

- Media assessment for new brand or new production lot
- Stock management and keeping of dehydrated media
- Production batch testing
  - pH
  - Sterility
  - Color and texture
  - Performance growth
  - Volume test

## 1. Internal quality control

### 1.6 Culture media quality and water supplied

- Media storage and shelf life
- Reagent-grade water quality
  - pH
  - Conductivity
  - Total metal
  - Residual chlorine
  - Total bacteria

## 1. Internal quality control

### 1.7 Glassware and laboratory apparatus

- Detergent and inhibitory residues
- Volume verification (Pipette, cylinder, micropipette, auto-dispenser)
- Sterility (For sterilized materials)

## 1. Internal quality control

### 1.8 Sample handling

- Sample identification
- Sample receiving and registration
- Sample storage
- Sample retention
- Sample disposal

## 1. Internal quality control

### 1.9 Quality control of test

- Positive and negative control
- Sterility control
- Replicate test

## 2. External quality assessment (Proficiency testing)

- Laboratory should regularly participate in proficiency testing to conduct performance evaluation studies among all participant laboratories.
- Results of such studies indicate the quality of analysis of each laboratory as compared to group performance

# **Quality assurance of temperature controlled equipments**

## The main temperature controlled equipments in microbiological laboratories

- Incubator
- Water bath
- Autoclave

# Three requirements for equipment handling

## Three requirements for equipment handling

### 1. Equipment maintenance

- Preventive maintenance plan by the rate of use
- Avoidance of cross-contamination
  - Clean and disinfect internal surfaces

Three requirements for equipment handling

## 2. Equipment calibration

- Calibration plan
- Uncertainty of measurement
- Temperature stability
- Temperature uniformity

Three requirements for equipment handling

### 3. Performance verification

- Maximum permissible error criteria (MPE)
- Time required to achieve equilibrium condition
- Monitor of the operating temperature with thermocouple or thermo data logger
  - Daily check
  - Each use check
  - Intermediate check

Three requirements for equipment handling

### 3. Performance verification

- Corrective action for non-conformity performance
- Validation after significant repair or modification

**Case study of  
temperature controlled equipment  
validation**

# Incubator

Installed place : Bacteria lab

Code No. : B-IN-09

Objective : For pathogenic bacteria growth

Temperature requirement :  $37^{\circ}\text{c}$


MPE :

Temperature uncertainty :  $\pm 1^{\circ}\text{c}$

Temperature uniformity :  $\pm 0.5^{\circ}\text{c}$

Temperature stability :  $\pm 0.5^{\circ}\text{c}$

# Result of calibration Incubator



**Result of Calibration :-** ( \* ) Without Adjustment


**Function of UUC\* :** Temperature Source

Calibration Point ( °C )	UUC* Setting ( °C )	UUC* Reading ( °C )	Average* Standard Reading ( °C )								
			Position								
			1	2	3	4	5	6	7	8	9 (ref.)
37.0	37.0	37.0	37.051	37.020	36.888	36.882	37.012	37.008	36.855	36.916	37.037

Calibration Point ( °C )	Temperature stability ( ± °C )	Temperature uniformity ( °C )	Overall Variation ( °C )	Uncertainty of measurement ( ± °C )	Coverage Factor k
37.0	0.031	0.20	0.26	0.31	2



# Monitor of operating temperature (Daily check)



# Water bath

Installed place : Bacteria lab

Code No. : B-WB-04

Objective : For selective growth of *E.coli* at 44.5°C  
and *Clostridium perfringens* at 46°C

Temperature requirement : 44.5°C and 46°C

MPE :

Temperature uncertainty at 44.5°C:  $\pm 0.2^\circ\text{C}$


Temperature uncertainty at 46°C:  $\pm 0.5^\circ\text{C}$

Temperature uniformity :  $\pm 0.5^\circ\text{C}$

Temperature stability :  $\pm 0.5^\circ\text{C}$

# Result of calibration Water bath

**Equipment** : Water Bath  
**Manufacturer** : Lauda  
**Model** : E125  
**Serial No.** : LCB0697-06-0309  
**ID No./Tag No.** : B-WB-04  
**Date Receipt** : 09-Jan-11  
**Date Calibrated** : 09-Jan-11  
**Ambient Temperature** : Start record 26 °C , Stop record 24 °C  
**Relative Humidity** : Start record 44 %RH , Stop record 42 %RH  
**Calibrated by** : Mr. Chaiya Engchoun



# Result of calibration Water bath

Table 1

Calibration Temperature (°C)	Setting Temperature (°C)	Indicating Temperature (°C)	Measured Stability (°C)	Measured Uniformity (°C)	Overall Variation (°C)
44.5	44.5	44.5	0.00	0.03	0.031

Table 2


Calibration Temperature (°C)	Standard Reading (°C), Probe No. 5 is Reference Probe					Uncertainty (±°C)
	No. 1	No. 2	No. 3	No. 4	No. 5	
44.5	44.48	44.50	44.49	44.48	44.50	0.14

# Verification of water bath performance

**Instrument Name:** Waterbath  
**Model:** E125  
**Serial No.:** LCB0697-06-0309  
**ID No.:** B-WB-04  
**Manufacturer:** Luada  
**Calibration Date:** January 9, 2011  
**Calibration by:** MIT  
**Temp Accept:**  $44.5 \pm 0.2^\circ\text{C}$  (44.3-44.7)

Verification parameter	Position1	Position2	Position3	Position4	Position5
UUC reading temp.	44.5	44.5	44.5	44.5	44.5
Reference reading temp.	44.48	44.50	44.49	44.48	44.50
Error	0.02	0.00	0.01	0.02	0.00
Uncertainty $\pm$	0.14	0.14	0.14	0.14	0.14
UUC reading - Error - Unc.	44.34	44.36	44.35	44.34	44.36
UUC reading - Error + Unc.	44.62	44.64	44.63	44.62	44.64
Maximum permissible error $44.5 \pm 0.2^\circ\text{C}$	44.3 - 44.7 $^\circ\text{C}$				
Result of verification	passed	passed	passed	passed	passed

# Monitor of operating temperature (In use check)



# Autoclave

Installed place : Sterilizing room

Code No. : B-AU-05

Objective : For culture media sterilization

Temperature requirement : 121°C

MPE :


Temperature uncertainty :  $\pm 3^{\circ}\text{C}$

# Result of calibration Autoclave

**Equipment :** Autoclave  
**Model :** HV-85  
**Serial No. :** 30406120400  
**ID No. :** B-AU-05  
**Manufacturer :** Hirayama  
**Received Order :** 11 January 2011  
**Condition As-Received :** Used Item  
**Calibrated Date :** 13 January 2011  
**Reference :** 101-304-1 OC  
**Result of Calibration :-** ( \* ) Without Adjustment

Cert. No.: 11TM194

Page.: 3 of 3



Operating parameter Set : Temperature = 121 °C

Sterilization period = 15 minutes

UUC* Setting ( °C )	UUC* Reading ( °C )	Position	Average* Standard Reading ( °C )	Stability ( ± °C )	Pressure Reading MPa	Uncertainty of measurement ( ± °C )	Coverage Factor k
121	121	1	121.229	0.13	0.12	0.86	2
		2	121.070				
		3	121.151				

# Verification of sterilization


## Result of calibration

UUC*	UUC*	Average* Standard Reading (°C)			Stability	Pressure Reading	Measurement Uncertainty
<u>Setting</u>	<u>Reading</u>	Position					
(°C)	(°C)	1	2	3	( +/- °C )	Mpa	( +/-°C )
121	121	121.229	121.070	121.151	0.13	0.12	0.86

Verification parameter	Position1	Position2	Position3
UUC reading temp.	121.0	121.0	121.0
Reference reading temp.	121.229	121.070	121.151
Error	-0.229	-0.070	-0.151
Uncertainty ±	0.86	0.86	0.86
UUC reading - Error - Unc.	-121.631	-121.79	-121.709
UUC reading - Error + Unc.	122.089	121.930	122.011
Maximum permissible error 121±3°C	118 - 124°C		
Result of verification	passed	passed	passed

## Monitor of sterilizing performance (In use check)

- Physical check
  - Maximum thermometer
  - Thermo data logger
- Chemical check
  - Chemical indicator strip
- Biological check
  - Bio-indicator  
*(Bacillus stearothermophilus)*



**Thank You**